

MAURITIAN
STANDARD

MS ISO
17123-1:2014

First Edition
2019-04-27

**Optics and optical instruments —
Field procedures for testing geodetic
and surveying instruments —**

**Part 1:
Theory**

ICS: 17.180.30



**Mauritius Standards Bureau
Moka**

National foreword

This Mauritian Standard is identical with the International Standard ISO 17123-1:2014, *Optics and optical instruments — Field procedures for testing geodetic and surveying instruments — Part 1: Theory*

It has been adopted as a national standard upon the recommendation of the **Metrology Standards Committee**.

It was approved by the Standards Council on 28 March 2019 and was notified in the Government Gazette on 27 April 2019.

For the purpose of this standard the following changes should be made:

- (i) the words 'International Standard' should be replaced by 'Mauritian Standard'
- (ii) the 'decimal comma' should be replaced by 'decimal point'.

General Notice No 881 of 2019



COPYRIGHT PROTECTED DOCUMENT

© MSB 2019

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission from Mauritius Standards Bureau at the address below

*Mauritius Standards Bureau
Villa Road
Moka
Mauritius*

Telephone + (230) 433 3648
Fax + (230) 433 5051/ 433 5150

E-mail msb@intnet.mu
Website <http://msb.intnet.mu>

Contents

	Page
Foreword	iv
Introduction	v
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
3.1 General metrological terms	1
3.2 Terms specific to this part of ISO 17123	3
3.3 The term “uncertainty”	5
3.4 Symbols	6
4 Evaluating uncertainty of measurement	8
4.1 General	8
4.2 Type A evaluation of standard uncertainty	9
4.3 Type B evaluation of standard uncertainty	18
4.4 Law of propagation of uncertainty and combined standard uncertainty	19
4.5 Expanded uncertainty	21
5 Reporting uncertainty	22
6 Summarized concept of uncertainty evaluation	22
7 Statistical tests	23
7.1 General	23
7.2 Question a): is the experimental standard deviation, s , smaller than or equal to a given value σ ?	23
7.3 Question b): Do two samples belong to the same population?	24
7.4 Question c) [respectively question d]): Testing the significance of a parameter y_k	24
Annex A (informative) Probability distributions	26
Annex B (normative) χ^2 distribution, Fisher’s distribution and Student’s t-distribution	27
Annex C (informative) Examples	28
Bibliography	39

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the WTO principles in the Technical Barriers to Trade (TBT) see the following URL: Foreword - Supplementary information

The committee responsible for this document is ISO/TC 172, *Optics and photonics*, Subcommittee SC 6, *Geodetic and surveying instruments*.

This third edition cancels and replaces the second edition (ISO 17123-1:2010).

ISO 17123 consists of the following parts, under the general title *Optics and optical instruments — Field procedures for testing geodetic and surveying instruments*:

- *Part 1: Theory*
- *Part 2: Levels*
- *Part 3: Theodolites*
- *Part 4: Electro-optical distance meters (EDM measurements to reflectors)*
- *Part 5: Total stations*
- *Part 6: Rotating lasers*
- *Part 7: Optical plumbing instruments*
- *Part 8: GNSS field measurement systems in real-time kinematic (RTK)*

Introduction

This part of ISO 17123 specifies field procedures for adoption when determining and evaluating the uncertainty of measurement results obtained by geodetic instruments and their ancillary equipment, when used in building and surveying measuring tasks. Primarily, these tests are intended to be field verifications of suitability of a particular instrument for the immediate task. They are not proposed as tests for acceptance or performance evaluations that are more comprehensive in nature.

The definition and concept of uncertainty as a quantitative attribute to the final result of measurement was developed mainly in the last two decades, even though error analysis has already long been a part of all measurement sciences. After several stages, the CIPM (Comité Internationale des Poids et Mesures) referred the task of developing a detailed guide to ISO. Under the responsibility of the ISO Technical Advisory Group on Metrology (TAG 4), and in conjunction with six worldwide metrology organizations, a guidance document on the expression of measurement uncertainty was compiled with the objective of providing rules for use within standardization, calibration, laboratory, accreditation and metrology services. ISO/IEC Guide 98-3 was first published as an International Standard (ISO document) in 1995.

With the introduction of uncertainty in measurement in ISO 17123 (all parts), it is intended to finally provide a uniform, quantitative expression of measurement uncertainty in geodetic metrology with the aim of meeting the requirements of customers.

ISO 17123 (all parts) provides not only a means of evaluating the precision (experimental standard deviation) of an instrument, but also a tool for defining an uncertainty budget, which allows for the summation of all uncertainty components, whether they are random or systematic, to a representative measure of accuracy, i.e. the combined standard uncertainty.

ISO 17123 (all parts) therefore provides, for defining for each instrument investigated by the procedures, a proposal for additional, typical influence quantities, which can be expected during practical use. The customer can estimate, for a specific application, the relevant standard uncertainty components in order to derive and state the uncertainty of the measuring result.

PREVIEW

Optics and optical instruments — Field procedures for testing geodetic and surveying instruments —

Part 1: Theory

1 Scope

This part of ISO 17123 gives guidance to provide general rules for evaluating and expressing uncertainty in measurement for use in the specifications of the test procedures of ISO 17123-2, ISO 17123-3, ISO 17123-4, ISO 17123-5, ISO 17123-6, ISO 17123-7 and ISO 17123-8.

ISO 17123-2, ISO 17123-3, ISO 17123-4, ISO 17123-5, ISO 17123-6, ISO 17123-7 and ISO 17123-8 specify only field test procedures for geodetic instruments without ensuring traceability in accordance with ISO/IEC Guide 99. For the purpose of ensuring traceability, it is intended that the instrument be calibrated in the testing laboratory in advance.

This part of ISO 17123 is a simplified version based on ISO/IEC Guide 98-3 and deals with the problems related to the specific field of geodetic test measurements.

2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO/IEC Guide 99, *International vocabulary of metrology — Basic and general concepts and associated terms (VIM)*

3 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO/IEC Guide 99 and the following apply.

3.1 General metrological terms

3.1.1

(measurable) quantity

property of a phenomenon, body or substance, where the property has a magnitude that can be expressed as a number and a reference

EXAMPLE 1 Quantities in a general sense: length, time, temperature.

EXAMPLE 2 Quantities in a particular sense: length of a rod.

3.1.2

value

value of a quantity

quantity value

number and reference together expressing the magnitude of a quantity

EXAMPLE Length of a rod: 3,24 m.